

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code	86048				
Lot No.	874		2023-09		2022-09

## REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	874	2023-09	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	874	2023-09	1
C3331	CONTROL SERUM CARDIOLIPIN-M	874	2023-09	1

## CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	365	$\geq 100$
Control Serum (MPLU/mL)	53,3	40,9-93,8

## CALIBRATOR

3	3	6	8	7	4	6	6	1	6	2	7	3	5	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature .....



## CONTROL SERUM

3	3	6	8	7	4	6	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 13.12.22

## WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.