

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME SM - CHORUS

Code	86016				
Lot No.	890		2024-02		2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3161	SM	890	2024-02	6X6
B3161	CALIBRATORE SM	890	2024-02	1
C3161	CONTROL SERUM SM	890	2024-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	477	≥ 100
Control Serum (AU/mL)	44,7	27,9-64,7

CALIBRATOR

1	9	2	8	9	0	7	2	1	7	2	9	4	3	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

1	9	2	8	9	0	7	1	6	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 14/12/22

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.