

**DIESSE DIAGNOSTICA SENESE SpA**  
 53035 Monteriggioni (SI) – Italia – Via delle Rose 10  
 Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione  
 20144 Milano (Italy) Via A. Solari 19 sc. 6  
 Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

**PRODUCT NAME:** VES-TEC

**Code** 10201/A

**Lot No** 4822



2024-02



2022-11

**Reference** IO-09/89

**ON LINE QUALITY CONTROLS** (model CDP not enclosed)

**QC Passed**

- sodium citrate solution concentration ☒
- sodium citrate volume ☒
- vacuum level (\*) ☒
- correct placement of the cap ☒
- no extra body inside the product ☒

(\*) Only for VACU-CODE and VACU-TEC tubes

**PACKAGING QUALITY CONTROL** (model CDP not enclosed)

**QC Passed**

- control of the number of the product for each pack ☒

**QUALITY CONTROLS ON FINISHED PRODUCT** (model CDP not enclosed)

**QC Passed**

- pack integrity ☒
- data printed on the product label ☒

**QC Passed**

N/A

**Irradiation treatment Certificate n° date**



**QC Passed**

N/A

**Gamma rays treatment process in compliance with the specifications of the supply Agreement**



Document number:

ODV N.A.

*This batch has been manufactured and checked according to quality requirements.  
 Hereby, it is released to distribution.*

**APPROVED**

RESPONSIBLE HEAD, Quality Control Reagents.....

**DATE** 02/12/2022