

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME SS-B - CHORUS

Code 86024

Lot No. 888



2023-11



2022-11

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3201	SS-B	888	2023-11	6X6
B3201	CALIBRATORE SS-B	888	2023-11	1
C3201	CONTROL SERUM SS-B	888	2023-11	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	402	$\geq 100$
Control Serum (AU/mL)	50,4	34,3-68,9


### CALIBRATOR

1	9	4	8	8	8	6	3	1	6	3	0	3	8	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



### CONTROL SERUM

1	9	4	8	8	8	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 19.12.2022

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.