

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME RF-M - CHORUS

Code 86040

Lot No. 879



2023-10



2022-10

REAGENTS

Label	Description	Lot.	Exp.	NO
D7451	RF-M	879	2023-10	6X6
B7451	CALIBRATORE RF-M	879	2023-10	1
C7451	CONTROL SERUM RF-M	879	2023-10	1

CONTROLS

	Found results	Expected results
--	---------------	------------------


Calibrator (mOD)	472	≥ 100
Control Serum (IU/mL)	19,8	16,1-26,3

CALIBRATOR

1	6	6	8	7	9	6	3	4	1	4	3	6	5	5	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒

Operator Signature *SB*




CONTROL SERUM

1	6	6	8	7	9	6	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒

Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *SB*

DATE 23.12.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.