

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME dsDNA-G - CHORUS

Code 86032

Lot No. 887



2023-11



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3281	dsDNA-G	887	2023-11	6X6
B3281	CALIBRATORE dsDNA-G	887	2023-11	1
C3281	CONTROL SERUM dsDNA-G	887	2023-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	420	≥ 100
Control Serum (IU/mL)	69,0	40,6-100,0

CALIBRATOR

2	5	1	8	8	7	6	4	1	8	2	8	4	6	3	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	5	1	8	8	7	6	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 23.12.2022

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.