

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GBM - CHORUS

Code	86070				
Lot No.	896		2023-12		2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3551	GBM	896	2023-12	6X6
B3551	CALIBRATORE GBM	896	2023-12	1
C3551	CONTROL SERUM GBM	896	2023-12	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	611	≥ 100
Control Serum (AU/mL)	32,2	26,7-55,7

CALIBRATOR

2	6	2	8	9	6	6	4	1	9	3	3	4	2	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

2	6	2	8	9	6	6	1	7	3	1	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 05/01/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.