



DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

## QUALITY CONTROL CERTIFICATE

**KIT NAME** CHORUS VITAMIN EXTRACTION

Code 86500

Lot No. 001

2025-02

2022-12

## REAGENTS

Label	Description	Lot.	Exp	No
B4580	Releasing Agent	001	2025-06	1
C4580	Stabilizing Agent	001	2025-02	1
D4580	Neutralizing Agent	001	2025-06	1

## CONTROLS

	Found results	Expected results
pH Releasing Agent	13,4	> 13,0
pH Stabilizing Agent	<2,0	< 2,0
pH Neutralizing Agent	4,26	4,0-4,5

*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE

06.02.2023