

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME CARDIOLIPIN-M - CHORUS

Code 86048

Lot No. 896

2023-12

2022-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3331	CARDIOLIPIN-M	896	2023-12	6X6
B3331	CALIBRATORE CARDIOLIPIN-M	896	2023-12	1
C3331	CONTROL SERUM CARDIOLIPIN-M	896	2023-12	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	415	$\geq 100$
Control Serum (MPLU/mL)	62,0	36,7-90,8

### CALIBRATOR

3	3	6	8	9	6	6	4	1	7	2	6	3	7	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature .....



### CONTROL SERUM

3	3	6	8	9	6	6	1	9	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 18.01.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.