

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc., 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

Code 86066

Lot No. 888



2023-11



2022-11

REAGENTS

Label	Description	Lot.	Exp.	NO
D3512	PR3	888	2023-11	6X6
B3512	CALIBRATORE PR3	888	2023-11	1
C3512	CONTROL SERUM PR3	888	2023-11	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	401	≥ 100
Control Serum (AU/mL)	67,5	47,1-87,4

CALIBRATOR


2	5	2	8	8	8	6	5	1	3	2	7	3	3	4	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	5	2	8	8	8	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.