

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME ADENOVIRUS IgA - CHORUS

Code 81198

Lot No. 900



2024-01



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D7611	Adenovirus IgA	900	2024-01	6X6
B7611	CALIBRATORE Adenovirus IgA	900	2024-01	1
C7611	CONTROL SERUM Adenovirus IgA	900	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	394	≥ 100
Control Serum (Ndx)	1,6	1,2-2,3
Negative Control (Ndx)	0,1	$<0,7$


CALIBRATOR

0	7	5	9	0	0	6	0	0	0	0	0	5	7	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *fe*




CONTROL SERUM

0	7	5	9	0	0	6	1	2	2	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *fe*



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *fe*

DATE *18.01.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.