

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME MPO - CHORUS

|         |       |  |         |  |         |
|---------|-------|--|---------|--|---------|
| Code    | 86068 |  |         |  |         |
| Lot No. | 898   |  | 2023-12 |  | 2022-12 |

### REAGENTS

| Label | Description       | Lot. | Exp.    | NO  |
|-------|-------------------|------|---------|-----|
| D3531 | MPO               | 898  | 2023-12 | 6X6 |
| B3531 | CALIBRATORE MPO   | 898  | 2023-12 | 1   |
| C3531 | CONTROL SERUM MPO | 898  | 2023-12 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                       |      |            |
|-----------------------|------|------------|
| Calibrator (mOD)      | 411  | $\geq 100$ |
| Control Serum (AU/mL) | 56,7 | 37,5-93,3  |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 3 | 8 | 9 | 8 | 6 | 1 | 1 | 7 | 2 | 5 | 3 | 7 | 3 | 7 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature *SB*



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 2 | 5 | 3 | 8 | 9 | 8 | 6 | 1 | 9 | 3 | 5 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature *SB*



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *SB*

DATE 20/01/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.