

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Vitamin D Total- CHORUS

Code 86900

Lot No. 896



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D9702	25OH Vitamin D Total	896	2023-12	6x6
B9703	CALIBRATOR 25OH Vitamin D Total	896	2024-12	1
C9703	CONTROL SERUM 25OH Vitamin D Total	896	2024-12	4

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	424	≥ 100
CONTROL SERUM (ng/ml)	33,7	28,6-36,4


CALIBRATOR

8	3	5	8	9	6	6	4	4	0	4	2	4	5	6	3	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 




CONTROL SERUM

8	3	5	8	9	6	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... 

DATE 30/01/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.