

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F17 HAZELNUT

Code 88000-F17

Lot No. 904



2024-01



2023-01

### REAGENTS

| Label | Description                      | Lot. | Exp.    | NO  |
|-------|----------------------------------|------|---------|-----|
| D9841 | 88000-F17 HAZELNUT               | 904  | 2024-01 | 2x6 |
| B9840 | CALIBRATORE 88000-F17 HAZELNUT   | 904  | 2024-02 | 1   |
| C9840 | CONTROL SERUM 88000-F17 HAZELNUT | 904  | 2024-02 | 1   |

### CONTROLS

|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                      |      |            |
|----------------------|------|------------|
| Calibrator (mOD)     | 260  | $\geq 100$ |
| Control Serum (KU/L) | 14,0 | 7,9-17,6   |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 3 | 9 | 9 | 0 | 4 | 6 | 5 | 1 | 7 | 5 | 3 | 2 | 3 | 9 | 5 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature .....

SB



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 6 | 3 | 9 | 9 | 0 | 4 | 6 | 3 | 0 | 5 | 6 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|


Label checked with barcode reader

YES



Operator Signature .....

SB



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

SB

DATE 30/01/2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.