

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-587111 Fax +39-0577-318690



Pag 1/1
Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-W5 WORMWOOD

Code	88000-W5		
Lot No.	904	2024-01	2023-02

REAGENTS

Label	Description	Lot	Exp	No
D4991	88000-W5 WORMWOOD	904	2024-01	2X6
B4990	CALIBRATORE 88000-W5 WORMWOOD	904	2024-02	1
C4990	CONTROL SERUM 88000-W5 WORMWOOD	904	2024-02	1

CONTROLS

	Found results	Expected result
--	---------------	-----------------

Calibrator (mOD)	260	≥ 100
Control Serum (KU/L)	14,0	7,9-17,6

CALIBRATOR


6	2	6	9	0	4	6	5	1	7	5	3	2	3	9	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *PA*



CONTROL SERUM


6	2	6	9	0	4	6	3	0	5	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *PA*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

DATE

13.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.