

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TRAB - CHORUS

Code 86112

Lot No. 900



2023-10



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D4041	TRAB	900	2023-10	6X6
B4040	CALIBRATORE TRAB	900	2024-01	1
C4040	CONTROL SERUM TRAB	900	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	297	≥ 100
Control Serum (U/L)	2,4	1,2-5,0


CALIBRATOR

9	4	7	9	0	0	5	6	2	5	4	0	3	5	4	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

9	4	7	9	0	0	5	2	0	3	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *10.02.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.