

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME SS-A - CHORUS

Code 86018

Lot No. 890



2024-02



2022-11

### REAGENTS

Label	Description	Lot.	Exp.	NO
D3181	SS-A	890	2024-02	6X6
B3181	CALIBRATORE SS-A	890	2024-02	1
C3181	CONTROL SERUM SS-A	890	2024-02	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	$\geq 100$
Control Serum (AU/mL)	35,9	22,5-50,0

### CALIBRATOR


1	9	3	8	9	0	7	1	2	0	3	2	4	6	5	1	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ..... *FB* .....



### CONTROL SERUM


1	9	3	8	9	0	7	1	5	2	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ..... *FB* .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *FB* .....

DATE *09.02.2023*

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.