DIESSE DIAGNOSTICA SENESE SPA

53035 Monteriggioni (SI), Italy, Strada dei Laghi 39 Tel. +39 0577319554 FAX +39 0577319020



Sede Legale 20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE SS-B - CHORUS

Code 86024

Lot No. 890

2023-11

2022-11

REAGENTS

KIT NAME

Label	Description	Lot,	Exp.	NO
D3201	SS-B	890	2023-11	6X6
B3201	CALIBRATORE SS-B	890	2023-11	1
C3201	CONTROL SERUM SS-B	890	2023-11	1

CONTROLS

Found results	Expected results

Calibrator (mOD)	360	≥ 100
Control Serum (AU/mL)	47,0	30,0-62,0

CALIBRATOR

1 9 4 8 9 0 6 3 2 0 3 5 3 6 5 5 0 0

Label checked with barcode reader

YES X

Operator Signature



CONTROL SERUM

1 9 4 8 9 0 6 2 0 3 6 0 0 0 0 0 1

Label checked with barcode reader

YES X

Operator Signature ...



This batch has been manufactured and checked according to quality requirements. Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents....

DATEUL, 02. 623

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.