

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME DIESSE W.W.-DAT

Code 26010

LOT.NO 153



2024-05



2023-02

REAGENTS

CODE	LABEL	DESCRIPTION	LOT	EXP	NO
V301	B1970	S. TYPHI O	157	2024-05	1
V302	E1970	S. TYPHI H	156	2024-05	1
V303	F1970	S. PARATYPHI A-O	143	2024-05	1
V304	G1970	S. PARATYPHI A-H	149	2024-05	1
V305	H1970	S. PARATYPHI B-O	144	2024-05	1
V306	I1970	S. PARATYPHI B-H	143	2024-05	1
V307	L1970	BRUCELLA	148	2024-05	1
PF21250/I	C1970	CONTROLLO POSITIVO	171	2024-05	1
PF02102/DAT	D1970	CONTROLLO NEGATIVO	136	2024-05	1

CONTROLS

FOUND RESULTS		EXPECTED RESULTS
CONTROLLO POSITIVO	Title	Title
S. TYPHI O	1/160	1/160 +/- 1 Dil
S. TYPHI H	1/160	1/160 +/- 1 Dil
S. PARATYPHI A-O	1/320	1/320 +/- 1 Dil
S. PARATYPHI A-H	1/160	1/160 +/- 1 Dil
S. PARATYPHI B-O	1/320	1/640 +/- 1 Dil
S. PARATYPHI B-H	1/320	1/320 +/- 1 Dil
BRUCELLA	1/320	1/640 +/- 1 Dil

Evaluation of the method according to the instruction for use.

This batch has been manufactured and checked according to quality requirements.

Hereby, it is released to distribution.

APPROVED BY: RESPONSIBLE HEAD, Quality Control Reagents

DATE:

16.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.