

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SIENA) – Italy – Strada dei Laghi, 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: AMA-M2 - CHORUS

Code	86084			
Lot No.	888		2024-02	
				2022-12

REAGENTS

Label	Description	Lot.	Exp	No
D3651	AMA-M2	888	2024-02	6X6
B3651	CALIBRATORE AMA-M2	888	2024-02	1
C3651	CONTROL SERUM AMA-M2	888	2024-02	1

CONTROLS


Found results	Expected results
---------------	------------------

Calibrator (mOD)	505	≥ 100
Control Serum (AU/mL)	35,3	32,5-100,0

CALIBRATOR

7	0	5	8	8	8	7	4	2	4	3	2	4	2	4	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature *f*



CONTROL SERUM

7	0	5	8	8	8	7	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *f*



The test has been carried out in accordance with the package insert

TESTED BY: Longo Elvira *f*
APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Longo*
DATE 16.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.