

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CHORUS PHOSPHOLIPID SCREEN IgG

Code 86106

Lot No. 895



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D4010	Phospholipid Screen IgG	895	2023-12	6X6
B4010	CALIBRATORE Phospholipid Screen IgG	895	2023-12	1
C4010	CONTROL SERUM Phospholipid Screen IgG	895	2023-12	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	467	≥ 100
Control (AU/ml)	32,8	25,7-52,0

CALIBRATOR


9	7	4	8	9	5	6	6	1	6	3	0	3	7	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



CONTROL SERUM


9	7	4	8	9	5	6	2	0	3	6	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 13.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.