

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada del Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

DIESSE
DIESSE

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code 86601

Lot No. 909



2024-02



2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D9970	CALPROTECTIN	909	2024-02	2X6
B9970	CALIBRATORE CALPROTECTIN	909	2024-05	1
C9970	CONTROL CALPROTECTIN	909	2024-05	1
G9502	SAMPLE DILUENT	909	2024-12	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	290	≥ 100
Control Serum (mg/Kg)	484,8	272,2-700,0

CALIBRATOR


9 2 9 9 0 9 6 2 2 8 4 8 2 7 5 5 0 0

Label checked with barcode reader

YES



Operator Signature *RB*



CONTROL SERUM


9 2 9 9 0 9 6 2 6 4 8 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature *RB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *[Signature]*

DATE 24.02.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.