

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Epstein-Barr VCA IgG - CHORUS

Code	81055				
Lot No.	900		2024-04		2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D2341	Epstein-Barr VCA IgG	900	2024-04	6X6
B2342	CALIBRATORE Epstein-Barr VCA IgG	900	2024-04	1
C2341	CONTROL SERUM Epstein-Barr VCA IgG	900	2024-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	238	≥ 100
Control Serum (INDEX)	3,6	3,1-5,7
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	0	5	9	0	0	7	0	0	0	0	0	2	3	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

0	0	5	9	0	0	7	3	1	5	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 24.02.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.