

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-319554 Fax +39-0577-319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: LKM-1 - CHORUS

Code	86080				
Lot No.	910		2024-02		2023-02

REAGENTS

Label	Description	Lot.	Exp	No
D3641	LKM-1	910	2024-02	6X6
B3641	CALIBRATORE LKM-1	910	2024-02	1
C3641	CONTROL SERUM LKM-1	910	2024-02	1

CONTROLS

Found results	Expected results
---------------	------------------

Calibrator (mOD)	727	≥ 100
Control Serum (AU/mL)	33,3	18,9-53,6

CALIBRATOR


7	0	3	9	1	0	6	4	1	9	3	3	6	7	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ...



CONTROL SERUM

7	0	3	9	1	0	6	1	1	2	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature ...



The test has been carried out in accordance with the package insert

TESTED BY: Antonella Leo

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 03.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.