

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Strada dei Laghi 39
Tel +39-0577-319554 Fax +39-0577-319020



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME: 88000-W21 WALL PELLITORY W21

Code	88000-W21		
Lot No.	910	2024-02	2023-03

REAGENTS

Label	Description	Lot	Exp	No
D9961	88000-W21 WALL PELLITORY W21	910	2024-02	2X6
B9960	CALIBRATORE 88000-W21 WALL PELLITORY W21	910	2024-05	1
C9960	CONTROL SERUM 88000-W21 WALL PELLITORY W21	910	2024-05	1

CONTROLS

	Found results	Expected result
Calibrator (mOD)	341	≥ 100
Control Serum (KU/L)	8,2	6,7-15,6

CALIBRATOR


6	1	6	9	1	0	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



CONTROL SERUM

6	1	6	9	1	0	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *RB*



The test has been carried out in accordance with the package insert

TESTED BY: Simone Baiocchi

Simone Baiocchi

APPROVED

RESPONSIBLE HEAD, Quality Control Reagents

[Signature]

DATE

06-03-23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested by an FDA approved method and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.