

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

## QUALITY CONTROL CERTIFICATE

KIT NAME 88000-F44 STRAWBERRY

Code 88000-F44

Lot No. 910



2024-02



2023-03

### REAGENTS

Label	Description	Lot.	Exp.	NO
D5741	88000-F44 STRAWBERRY	910	2024-02	2x6
B5740	CALIBRATORE 88000-F44 STRAWBERRY	910	2024-05	1
C5740	CONTROL SERUM 88000-F44 STRAWBERRY	910	2024-05	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	341	$\geq 100$
Control Serum (KU/L)	8,2	6,7-15,6

### CALIBRATOR

6	4	0	9	1	0	6	5	2	0	6	2	2	4	7	8	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

BB



### CONTROL SERUM

6	4	0	9	1	0	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature

BB



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 03.03.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.