

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BORDETELLA PERTUSSIS TOXIN IgG - CHORUS

Code 81272

Lot No. 917



2024-03



2023-02

REAGENTS

Label	Description	Lot.	Exp	No
D2100	BORDETELLA PERTUSSIS TOXIN IgG	917	2024-03	6x6
B2100	CALIBRATORE BORDETELLA P.TOXIN IgG	917	2024-05	1
C2100	CONTROL SERUM BORDETELLA P.TOXIN IgG	917	2024-05	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	490	≥ 100
CONTROL SERUM (IU/ml)	176,3	133,3-350,0

CALIBRATOR

9	6	4	9	1	7	6	1	1	6	3	1	3	3	4	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *BS*



CONTROL SERUM


9	6	4	9	1	7	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *BS*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *BS*

DATE 07/03/2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.