

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME ANTI-CCP - CHORUS

Code	86094		
Lot No.	899	2023-12	2022-12

### REAGENTS

Label	Description	Lot.	Exp.	NO
D8671	ANTI-CCP	899	2023-12	6x6
B8671	CALIBRATORE ANTI-CCP	899	2023-12	1
C8671	CONTROL SERUM ANTI-CCP	899	2023-12	1


### CONTROLS

	Found results	Expected results
Calibrator (mOD)	487	≥ 100
Control Serum (AU/mL)	61,9	46,1-96,9

### CALIBRATOR

2	5	6	8	9	9	6	5	1	9	3	5	3	9	5	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



### CONTROL SERUM

2	5	6	8	9	9	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 13-03-2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.