

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 8

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Chorus Chlamydia trachomatis IgG

Code 81261

Lot No. 917



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D9640	Chorus Chlamydia trachomatis IgG	917	2024-03	6X6
B9640	CALIBRATOR Chlamydia trachomatis IgG	917	2024-03	1
C9640	POSITIVE CONTROL Chlamydia trachomatis IgG	917	2024-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	550	≥ 100
Positive Control (Index)	1,3	0,9-1,7

CALIBRATOR

2	3	3	9	1	7	6	0	0	0	0	0	7	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

2	3	3	9	1	7	6	0	9	1	7	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE *15-03-2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.