

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME HSV 1 SCREEN - CHORUS

Code 81018

Lot No. 909



2024-05



2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2191	HSV 1 SCREEN	909	2024-05	6x6
B2191	CALIBRATORE HSV 1 SCREEN	909	2024-05	1
C2191	CONTROL SERUM HSV 1 SCREEN	909	2024-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (MOD)	460	≥ 100
CONTROL SERUM (Index)	2,0	1,3-8,0
NEGATIVE CONTROL (Index)	0,6	$< 0,7$

CALIBRATOR

0	3	3	9	0	9	7	0	0	0	0	0	6	5	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


0	3	3	9	0	9	7	1	3	8	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 15-03-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent reagents should be handled with maximum attention.