

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME CALPROTECTIN K - CHORUS

Code	86601				
Lot No.	910		2024-02		2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D9970	CALPROTECTIN	910	2024-02	2X6
B9970	CALIBRATORE CALPROTECTIN	910	2024-05	1
C9970	CONTROL CALPROTECTIN	910	2024-05	1
C9502	SAMPLE DILUENT	910	2025-02	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	270	≥ 100
Control Serum (mg/Kg)	445,9	272,2-700,0

CALIBRATOR

9	2	9	9	1	0	6	2	2	8	4	8	2	7	5	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

9	2	9	9	1	0	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.

DATE 16/03/23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.