



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI) – Italia – Via delle Rose 10
Tel +39-0577-587111 Fax +39-0577-318690



Sede Legale e Amministrazione
20144 Milano (Italy) Via A. Solari 19 sc. 6
Tel. +39-02-4859121 r.a. Fax +39-02-48008530

Pag 1/1

PRODUCT NAME: VACU-TEC

| | |
|---|----------------|
| Code | 10208 |
| Lot No | 3072 |
|  | 2024-01 |
|  | 2023-03 |
| Reference IO-09/88 | |

ON LINE QUALITY CONTROLS (model CDP not enclosed)**QC Passed**

- sodium citrate solution concentration
- sodium citrate volume
- vacuum level (*)
- correct placement of the cap
- no extra body inside the product

☒
☒
☒
☒
☒
☒

(*) Only for VACU-CODE and VACU-TEC tubes

PACKAGING QUALITY CONTROL (model CDP not enclosed)**QC Passed**

- control of the number of the product for each pack

☒

QUALITY CONTROLS ON FINISHED PRODUCT (model CDP not enclosed)**QC Passed**

- pack integrity
- data printed on the product label

☒
☒

QC Passed**N/A****Irradiation treatment Certificate n° date**

☒

☒**QC Passed****N/A****Gamma rays treatment process in compliance with the specifications of the supply Agreement**

☒

☒

Document number:

ODV N.A.

*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVEDRESPONSIBLE HEAD, Quality Control Reagents.....**DATE** 15/03/2023