



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME HERPES SIMPLEX 1+2 IgG -CHORUS

Code 81020

Lot No. 897



2024-03



2023-02

REAGENTS

Label	Description	Lot.	Exp.	NO
D2211	Herpes simplex 1+2 IgG	897	2024-03	6X6
B2212	CALIBRATORE Herpes simplex 1+2 IgG	897	2024-05	1
C2211	CONTROL SERUM Herpes simplex 1+2 IgG	897	2024-05	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	380	≥ 100
Control Serum (INDEX)	2,0	1,6-3,0
Negative Control (INDEX)	0,0	$< 0,7$

CALIBRATOR

0	0	4	8	9	7	7	0	0	0	0	0	4	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*



CONTROL SERUM

0	0	4	8	9	7	7	1	6	3	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *MB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *MB*

DATE 21.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.