

QUALITY CONTROL CERTIFICATE

KIT NAME MPO - CHORUS

Code	86068		
Lot No.	917	 2024-03	 2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3531	MPO	917	2024-03	6X6
B3531	CALIBRATORE MPO	917	2024-03	1
C3531	CONTROL SERUM MPO	917	2024-03	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	515	> 100
Control Serum (AU/mL)	40,9	33,0-71,4

CALIBRATOR

2	5	3	9	1	7	6	3	1	5	2	5	4	5	3	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader YES ☒ Operator Signature 



CONTROL SERUM

2	5	3	9	1	7	6	1	6	2	9	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents. 

DATE 23-03-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.