

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME PR3 - CHORUS

Code 86066

Lot No. 902



2024-01



2023-01

REAGENTS

Label	Description	Lot.	Exp.	NO
D3512	PR3	902	2024-01	6X6
B3512	CALIBRATORE PR3	902	2024-01	1
C3512	CONTROL SERUM PR3	902	2024-01	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	> 100
Control Serum (AU/mL)	55,5	43,5-71,9

CALIBRATOR


2	5	2	9	0	2	6	7	1	2	3	2	3	3	5	9	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

2	5	2	9	0	2	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 22.03.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.