

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME Chorus Chlamydia trachomatis IgA

Code 81262

Lot No. 910



2024-02



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D9650	Chorus Chlamydia trachomatis IgA	910	2024-02	6X6
B9650	CALIBRATOR Chlamydia trachomatis IgA	910	2024-02	1
C9650	POSITIVE CONTROL Chlamydia trachomatis IgA	910	2024-02	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	460	≥ 100
Positive Control (Index)	1,5	1,0-1,8

CALIBRATOR

2	3	2	9	1	0	6	0	0	0	0	0	7	4	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



CONTROL SERUM

2	3	2	9	1	0	6	1	0	1	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *ABene*

DATE *31.03.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.