

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-G - CHORUS

Code 86056

Lot No. 917



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D3411	Gliadin-G	917	2024-03	6X6
B3411	CALIBRATORE Gliadin-G	917	2024-03	1
C3411	CONTROL SERUM Gliadin-G	917	2024-03	1

CONTROLS

	Found results	Expected results
CALIBRATOR (mOD)	638	≥ 100
CONTROL SERUM (AU/ml)	52,1	26,7-100,0

CALIBRATOR

5 0 2 9 1 7 6 4 1 6 2 3 5 7 4 0 0 0

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM

5 0 2 9 1 7 6 1 2 2 3 0 0 0 0 0 0 1

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 04.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.