

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME GLIADIN-A - CHORUS

Code 86054

Lot No. 897



2023-12



2022-12

REAGENTS

Label	Description	Lot.	Exp.	NO
D3391	Gliadin-A	897	2023-12	6X6
B3392	CALIBRATORE Gliadin-A	897	2024-03	1
C3392	CONTROL SERUM Gliadin-A	897	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	422	≥ 100
CONTROL SERUM (AU/ml)	91,8	58,8-147,1

CALIBRATOR


5	0	1	8	9	7	6	1	2	3	4	0	2	7	7	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



CONTROL SERUM


5	0	1	8	9	7	6	2	6	4	8	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 11.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.