

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

### QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA IgA - CHORUS

Code 81033

Lot No. 926



2024-04



2023-04

### REAGENTS

| Label | Description                  | Lot. | Exp.    | NO  |
|-------|------------------------------|------|---------|-----|
| D7551 | MYCOPLASMA IgA               | 926  | 2024-04 | 6X6 |
| B7551 | CALIBRATORE MYCOPLASMA IgA   | 926  | 2024-04 | 1   |
| C7551 | CONTROL SERUM MYCOPLASMA IgA | 926  | 2024-04 | 1   |

### CONTROLS


|  | Found results | Expected results |
|--|---------------|------------------|
|--|---------------|------------------|

|                          |       |            |
|--------------------------|-------|------------|
| Calibrator (mOD)         | 435   | $\geq 100$ |
| Control Serum (AU/mL)    | 40,6  | 27,5-54,1  |
| Negative Control (AU/mL) | <10,0 | <10        |

### CALIBRATOR

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 6 | 8 | 9 | 2 | 6 | 6 | 8 | 1 | 2 | 2 | 0 | 4 | 7 | 5 | 4 | 0 | 0 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature .....



### CONTROL SERUM

|   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |   |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|
| 0 | 6 | 8 | 9 | 2 | 6 | 6 | 1 | 5 | 2 | 8 | 0 | 0 | 0 | 0 | 0 | 0 | 1 |
|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|---|

Label checked with barcode reader YES ☒ Operator Signature .....



*This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 12.06.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.