

DIESE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME BRUCELLA IgM CHORUS

Code 81166

Lot No. 918



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D4120	Brucella IgM	918	2024-03	6X6
B4120	CALIBRATORE Brucella IgM	918	2024-03	1
C4120	Control Serum Brucella IgM	918	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	422	≥ 100
Control (Index)	1,9	1,4-2,5


CALIBRATOR

1	4	7	9	1	8	6	0	0	0	0	5	2	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



CONTROL SERUM

1	4	7	9	1	8	6	1	4	2	5	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *FB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents *FB*

DATE 18.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.