

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgG -CHORUS

Code 81010

Lot No. 918



2024-03



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2161	Cytomegalovirus IgG	918	2024-03	6x6
B2162	CALIBRATORE Cytomegalovirus IgG	918	2024-03	1
C2161	CONTROL SERUM Cytomegalovirus IgG	918	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	327	≥ 100
Control Serum (IU/mL)	11,0	5,7-15,8
Negative Control (IU/mL)	<0,5	<0,5

CALIBRATOR

0	0	1	9	1	8	6	5	2	7	4	1	2	4	6	5	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature *[Signature]*

9



CONTROL SERUM


0	0	1	9	1	8	6	2	9	5	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature *[Signature]*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *[Signature]*

DATE 30-06-2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.