



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

QUALITY CONTROL CERTIFICATE

KIT NAME INFLUENZA B IgA -CHORUS

Code 81195

Lot No. 920



2024-03



2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D7541	INFLUENZA B IgA	920	2024-03	6X6
B7541	CALIBRATORE INFLUENZA B IgA	920	2024-03	1
C7541	CONTROL SERUM INFLUENZA B IgA	920	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	420	≥ 100
Control Serum (Ndx)	2,6	1,8-3,4
Negative Control (Ndx)	0,5	$\leq 0,7$

CALIBRATOR

0	7	1	9	2	0	6	0	0	0	0	0	3	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



CONTROL SERUM

0	7	1	9	2	0	6	1	8	3	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---


Label checked with barcode reader

YES



Operator Signature

FB



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 20.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.