

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME TG Extended Range

Code	86110		
Lot No.	900	2024-01	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D4050	TG Extended Range	900	2024-01	6x6
B4050	CALIBRATOR TG Extended Range	900	2024-04	1
C4050	CONTROL SERUM TG Extended Range	900	2024-04	1

CONTROLS


	Found results	Expected results
--	---------------	------------------

CALIBRATOR (mOD)	380	≥ 100
CONTROL SERUM (ng/ml)	2,2	1,5-3,5

CALIBRATOR

4	1	8	9	0	0	6	7	1	3	2	3	3	7	5	4	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

4	1	8	9	0	0	6	1	9	3	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Decover*

DATE *21.06.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.