

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME TG - CHORUS

Code	86076		
Lot No.	926	2024-04	2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3613	TG	926	2024-04	6x6
B3613	CALIBRATORE TG	926	2024-04	1
C3613	CONTROL SERUM TG	926	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	408	≥ 100
Control Serum (ng/mL)	46,9	24,6-66,7

CALIBRATOR

4	2	0	9	2	6	6	3	1	9	3	2	3	3	5	6	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



CONTROL SERUM

4	2	0	9	2	6	6	2	1	4	0	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents

DATE 27.01.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.