

QUALITY CONTROL CERTIFICATE

KIT NAME TOXOPLASMA IgG - CHORUS

Code 81040

Lot No. 920

2024-03

2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2271	TOXOPLASMA IgG	920	2024-03	6X6
B2272	CALIBRATORE TOXOPLASMA IgG	920	2024-03	1
C2271	CONTROL SERUM TOXOPLASMA IgG	920	2024-03	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	342	≥ 100
Control Serum (IU/mL)	24,2	20,9-43,8
Negative Control (IU/mL)	<5,0	<5

CALIBRATOR

0	8	2	9	2	0	6	2	2	1	3	7	4	0	5	7	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *Beni*



CONTROL SERUM

0	8	2	9	2	0	6	1	8	3	3	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *Beni*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *Beni*

DATE 28.04.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.