



DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020

Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

QUALITY CONTROL CERTIFICATE

KIT NAME CYTOMEGALOVIRUS IgM CAPTURE - CHORUS

Code 81012

Lot No. 926



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D3920	CYTOMEGALOVIRUS IgM CAPTURE	926	2024-04	6X6
B3920	CALIBRATORE CYTOMEGALOVIRUS IgM CAPTURE	926	2024-04	1
C3920	CONTROL SERUM CYTOMEGALOVIRUS IgM CAPTURE	926	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	261	≥ 100
Control Serum (Index)	4,2	2,4-4,4


CALIBRATOR

9	7	1	9	2	6	6	0	0	0	0	0	3	0	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *HB*



CONTROL SERUM

9	7	1	9	2	6	6	2	4	4	4	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES ☒

Operator Signature *HB*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *HB*

DATE 03.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.