

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME MYCOPLASMA PNEUMONIAE IgG - CHORUS

Code 81034

Lot No. 919  2024-03  2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2251	MYCOPLASMA PNEUMONIAE IgG	919	2024-03	6x6
B2252	CALIBRATORE MYCOPLASMA PNEUMONIAE IgG	919	2024-03	1
C2251	CONTROL SERUM MYCOPLASMA PNEUMONIAE IgG	919	2024-03	1

CONTROLS

	Found results	Expected results
Calibrator (mOD)	510	≥ 100
Control Serum (UA/mL)	39,0	20,0-55,0
Negative Control (UA/mL)	<10,0	<10

CALIBRATOR



0 6 6 9 1 9 6 7 1 3 2 0 5 4 3 6 0 0

Label checked with barcode reader YES ☒ Operator Signature 

CONTROL SERUM

0 6 6 9 1 9 6 1 3 2 4 0 0 0 0 0 0 1

Label checked with barcode reader YES ☒ Operator Signature 



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... 

DATE 02.05.23

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.