

DIESSE DIAGNOSTICA SENESE SpA
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

QUALITY CONTROL CERTIFICATE

KIT NAME EPSTEIN-BARR EARLY ANTIGEN IgM

Code 81059

Lot No. 927



2024-04



2023-04

REAGENTS

Label	Description	Lot.	Exp.	NO
D2382	Epstein-Barr Early Antigen IgM	927	2024-04	6x6
B2382	CALIBRATORE Epstein-Barr Early Antigen IgM	927	2024-04	1
C2382	CONTROL SERUM Epstein-Barr Early Antigen IgM	927	2024-04	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	320	≥ 100
Control Serum (Index)	2,0	1,7-3,1
Negative Control (Index)	0,0	$< 0,7$


CALIBRATOR

0 2 9 9 2 7 6 0 0 0 0 0 4 2 0 0 0 0

Label checked with barcode reader

YES ☒

Operator Signature *calu*




CONTROL SERUM

0 2 9 9 2 7 6 1 7 3 1 0 0 0 0 0 0 1

Label checked with barcode reader

YES ☒

Operator Signature *calu*



*This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.*

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *calu*

DATE *06.05.2023*

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.