

DIESSE DIAGNOSTICA SENESE SpA  
53035 Monteriggioni (SI), Italy, Strada dei Laghi 39  
Tel. +39 0577319554 FAX +39 0577319020



Sede Legale  
20144 Milano (Italy) Via A. Solari 19 sc. 6

Pag 1/1

## QUALITY CONTROL CERTIFICATE

KIT NAME VARICELLA IgM CHORUS

Code 81079

Lot No. 910

2024-05



2023-03

### REAGENTS

Label	Description	Lot.	Exp.	NO
D2481	VARICELLA IgM	910	2024-05	6x6
B2481	CALIBRATORE VARICELLA IgM	910	2024-06	1
C2481	CONTROL SERUM VARICELLA IgM	910	2024-06	1

### CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	450	$\geq 100$
Control Serum (Index)	2,8	1,7-3,2
Negative Control (Index)	0,1	$< 0,7$

### CALIBRATOR


1	3	4	9	1	0	7	0	0	0	0	0	4	1	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



### CONTROL SERUM

1	3	4	9	1	0	7	1	7	3	2	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader

YES



Operator Signature .....



This batch has been manufactured and checked according to quality requirements.  
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents.....

DATE 05.05.2023

### WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.