

QUALITY CONTROL CERTIFICATE

KIT NAME MEASLES IgG CHORUS

Code	81072
Lot No.	918
	2024-06
	2023-03

REAGENTS

Label	Description	Lot.	Exp.	NO
D2431	MEASLES IgG	918	2024-06	6X6
B2432	CALIBRATORE MEASLES IgG	918	2024-06	1
C2431	CONTROL SERUM MEASLES IgG	918	2024-06	1

CONTROLS

	Found results	Expected results
--	---------------	------------------

Calibrator (mOD)	350	≥ 100
Control Serum (Index)	3,4	2,4-4,5
Negative Control (Index)	0,0	$< 0,7$

CALIBRATOR

0	0	9	9	1	8	7	0	0	0	0	0	2	9	0	0	0	0
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



CONTROL SERUM

0	0	9	9	1	8	7	2	4	4	5	0	0	0	0	0	0	1
---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---	---

Label checked with barcode reader YES ☒ Operator Signature *HB*



This batch has been manufactured and checked according to quality requirements.
Hereby, it is released to distribution.

APPROVED RESPONSIBLE HEAD, Quality Control Reagents..... *HB*

DATE 09.05.2023

WARNING: POTENTIAL BIOHAZARDOUS MATERIAL!

This kit may contain some reagents made with human serum or plasma. All serum or plasma used has been tested and found non-reactive for HIV-1/2, HCV and HBsAg. Because no method can offer complete assurance that HIV-1/2, HCV, HBsAg or other infectious agents are absent, reagents should be handled with maximum attention.